



Date – 31st March, 2016

ADDENDUM-II

Clarification(s) counter to queries for the supply of “Laboratory Furniture and Fumehood”

Tender No. IITGN/MSE/NON-CONS/2015-16/0482 Dated 8th March, 2016

1. Delivery period / Installation period:

It should always mentioned in Tender documents and not in order.

Further, please note that Our delivery period is 6-8 weeks from the date of clear Techno-commercial order. Is iT acceptable to you??

Response: You may mention your delivery period in the bid.

2. Warranty Period:

Till now neither any customer has asked or nor any supplier has given warranty for 36 month for Fume Hood. Please Check once again !!

Response: The Institute required a warranty for 36 months for Fumehood.

3. C-Frame:

No details are available in tender documents/table or drawing. Please check your reply again and clarify technical details of C-Frame as it is an important part of Furniture.

What is size of Frame...??, What is thickness of Frame...??, Size details of rectangular tube etc..??

Response: C-Frame should be manufactured from standard hollow metal sections, conforming to IS code 7138: 973 and sheet metal components should conform to IS code 513:1994. The rectangular pipe cross section should be 60mm X 30mm and 2mm thick. The frame dimensions should be according to the drawings in the tender document.

4. LHP Motors:

LHP is not a reputed in India. It is just local manufactures of motors Maharashtra. Please arrange to send us details of LHP as we would like to also quote for same.

Response: You may quote any standard make motor.

5. Airfoil 304:

Teflon coated SS304 is not at all chemical resistance. Please check your reply!!

Response: The institute require Teflon coated SS304, you may quote the same or better.

6. We request you to please arrange to extend due date for at least 3 to 4 weeks.

Response: The due date for submission of the bids remains unchanged. If the Institute at a later date feels required, the dates may be extended.

